



Streck Customer Notification

CALIBRATION STATEMENT

Prior to shipment, Temp-Chex thermometers are stored in a controlled environment. With proper care, the thermometer should maintain its accuracy. Due to use in diverse applications and handling, there is no precise way to predict how long calibration will be maintained. As with any thermometer, drift is a possibility and accuracy can be affected by shock, aging, temperature and contamination through daily use. Periodic recalibration or verification should be scheduled by each laboratory in accordance with its accrediting agency and/or established procedure. At a minimum, annual recalibration or verification from the date the thermometer is put "in use" is good lab practice. A space is provided at the top of the Certificate of Accuracy to record the "in use" date. The date of certification listed for this thermometer is for reference use only.

Please contact Streck Technical Service at 800-843-0912 or technicalservices@streck.com with any questions.

Temp-Chex Thermometer Model: _____

Serial Number: _____

In Use Date: _____